

**KANNUR UNIVERSITY**  
**(Purchase and Inventory Control Office A Section )**

PICO-/A2/14801/2024

20.12.2024

**Quotation for the Supply and Installation of Fume Hood for the Dept of Physics, Swami Anandatheertha Campus Payyannur**

Sealed Quotations superscribed as "**Supply, installation and of Fume Hood for the Department of Physics, Swami Anandatheertha campus, Payyannur**" are invited from authorized dealers having proven track record for the supply installation of Fume Hood, for the **Department of Physics, Swami Anandatheertha campus, Payyannur.**

Quotations with details of model, price, etc. should reach the **Registrar, Kannur University, Civil Station P.O, Kannur, 670002** on or before **06/01/2025, 2 PM** and Late Quotations will not be accepted. The Quotations will be opened at **3 PM** on the same day in the presence of such of the Quotationers or their authorized representatives who may be present at that time. The maximum period required for delivery of the articles should also be mentioned. The rate quoted should be inclusive of all taxes, installation charges and other charges. Quotation forms and General Conditions can be downloaded from University Website [www.kannuruniversity.ac.in](http://www.kannuruniversity.ac.in).

**QUOTATION SCHEUDLE**

QUOTATION NOTICE NO	No. PICO/A2/14801/2024
NAME OF WORK	Supply and installation and of Fume Hood at department of Physics
BID SUBMISSION END DATE AND TIME	<b>06.01.2025 at 2pm</b>
BID OPENING DATE AND TIME	<b>06.01.2025 at 3pm</b>
WARRANTY	<b>1 Year</b>

### TECHNICAL SPECIFICATION

Specifications	Quantity	Place of Supply, installation and commissioning.
<p>1. The hood should be fabricated from good-quality, waterproof marine plywood or CR Sheet. The outer surfaces should be covered with epoxy coating or FRP Covering as required.</p> <p>2. The inner surfaces should be covered with epoxy coating or FRP lining as required.</p> <p>3. The working table should be covered with a 20-mm-thick granite slab, provided with a water tap and stainless steel or ceramic sink as required.</p> <p>4. Should be provided with gas/Air Cock, additional power point, diffused fluorescent tube.</p> <p>5. The exhaust system should have a dynamically balanced Epoxy/FRP- coated impeller and a TEFC Industrial motor.</p> <p>6. The front glass shutter should have a counterweight arrangement for easy vertical sliding.</p> <p>7. The bottom stand should be provided with a cupboard and drawers.</p> <p>8. Should have Approx Ext. Dim: 126x75x240cm</p> <p>9. Should have Working Table Size: 120x60x60cm</p>	1 No.	Department of Physics , Swami Anandatheertha Campus, Payyannur

#### **Pricing**

The rate quoted should be inclusive of GST and all other charges, if any.

The basic rate, GST charges and all other charges, if any, should be furnished separately

#### **Delivery**

Fume Hood should be delivered and installed at the department of Physics , Swami Anandatheertha Campus, Edat (PO) , Payyannur, 670327. Authorized dealers having credible experience can only participate in the

Quotation. Item offered must be standard, branded & proven product and already available in the market.

**TERMS AND CONDITIONS**

1. The rate quoted should be inclusive of GST and all other charges, if any. The base rate, GST charges and other charges, if any, should be furnished separately.
2. Quotation should be valid for a minimum period of **three months**.
3. Details of Warranty Condition should be provided along with quotation.
4. Supplier shall deliver and Install the item at the locations mentioned above.
5. The Registrar, Kannur University reserves the right to accept or reject the quotation , without assigning any reason.

GSTIN 32AAAGK0152J1ZT

Contact : 04972715267

Sd/-

**REGISTRAR**

*Copy to : Web Manager/Notice Board*

